

Certificate of Compliance for medical and veterinary computed tomography X-ray apparatus



This certificate of compliance applies to the compliance testing of fixed and mobile medical and veterinary X-ray apparatus capable of computed tomography.

Part 1—Owner details

Please provide the details of the owner of the apparatus.

EPA Registration No:

--	--	--	--	--

Please refer to the EPA acknowledgment letter sent to the owner

Name

Address
(Street/Suburb)

State

Postcode

Part 2—Apparatus details

Please provide the details of the apparatus under test.

Make & model

Design

Medical

Veterinary

Fixed

Mobile

Generator S/N

Tube housing S/N

X-ray tube S/N

Room location

Address
(Street/Suburb)

State

Postcode

Part 3—Declaration

I hereby declare that I have compliance tested the apparatus detailed above and to the best of my knowledge the information provided on this certificate is true and correct.

Name (please print)

Date

Accreditation No.

Email address*

Signature

*A completed certificate is accepted by the EPA as being an electronic communication, under Division 2 section 8 of the *Electronic Transactions Act 2000*, if sent from an email address that the EPA can use as the method of identifying the originator of the electronic communication.

Please lodge the completed certificate by email at EPARPB.Compliance@epa.sa.gov.au

OFFICE USE ONLY

Owner No.

--	--	--	--	--

Reg. No.

--	--	--	--	--

BAL No.

--	--	--	--	--

Part 4—Assessment

The references below refer to the specified sections of *Test Protocol for Medical and Veterinary Computed Tomography X-ray Apparatus 2023*.

Ref.	Test Description	Compliance Verified		
Sec 8	Labelling of apparatus	Yes	No	N/A
Sec 9	Radiation area sign	Yes	No	N/A
Sec 10	Apparatus to be in good working order	Yes	No	N/A
Sec 11	Markings on X-ray generators and tube assemblies	Yes	No	N/A
Sec 12	Warning device	Yes	No	N/A
Sec 13	Exposure termination	Yes	No	N/A
Sec 14	Indicators of operation	Yes	No	N/A
Sec 15	Mechanical accuracy	Yes	No	N/A
Sec 16	Image quality	Yes	No	N/A
Sec 17	Reconstructed slice thickness	Yes	No	N/A
Sec 18	Computed tomography dose index in air	Yes	No	N/A
Sec 19	Computed tomography dose index volume in phantom	Yes	No	N/A

Part 5—Interpretation

Unless the contrary intention appears—

BAL means business activities licence

Compliance Verified means—

- (a) in the case where an accredited tester has indicated a 'Yes', compliance with the specified test has been verified;
- (b) in the case where an accredited tester has indicated a 'No', the apparatus is not compliant with the specified test or a verification of compliance is not possible;
- (c) in the case where an accredited tester has indicated a 'N/A', the test does not apply to the apparatus.

Sec means section of the specified test protocol

S/N means serial number